SADCAS

SOUTHERN AFRICAN DEVELOPMENT COMMUNITY ACCREDITATION SERVICE

CERTIFICATE OF ACCREDITATION

SEYCHELLES BUREAU OF STANDARDS (CALIBRATION)

Established by Section 3 of the Seychelles Bureau of Standards Act 1987

Facility Accreditation Number: CAL-8 001

is a SADCAS accredited Calibration Laboratory
provided that all SADCAS conditions are complied with

This certificate is valid as per the scope stated in the accompanying schedule of accreditation, Annexure "A", bearing the above accreditation number for

MASS METROLOGY

The facility is accredited in accordance with the recognized International Standard

ISO/IEC 17025:2017

The accreditation demonstrates technical competency for a defined scope and the operation of a laboratory quality management system

SADCAS is a subsidiarity organization of SADC. A memorandum of understanding between SADC and SADCAS serves as the basis for the recognition of SADCAS by SADC Member States as a multi-economy accreditation body

Eve Christine Gadzikwa
SADCAS Chief Executive Officer

Date of Renewal of Accreditation: 06 July 2025
Effective Date (Issue No: 1): 15 September 2025
Certificate Expires: 14 September 2030

SADCAS
SOUTHERN AFRICAN DEVELOPMENT COMMUNITY ACCREDITATION SERVICE

ANNEXURE A

SCHEDULE OF ACCREDITATION

MASS METROLOGY

Laboratory Accreditation Number: CAL-8 001 (ISO/IEC 17025:2017)

Permanent Address of Laboratory
Seychelles Bureau of Standards
Avenue d'Arhoa
Providence Industrial Estate
Mahe
Seychelles

Postal Address
P O Box 953
Mahe
Seychelles

Tel : +248 43 80400/43 80443
Cell : +248 258 2126
Fax : +248 437 3826
Email : hrichmond@sbs.sc

Technical Signatories : Mr H Richemond (All items)
Mr F M Duval (All items
Mr K Rose (All items)

Nominated Representative : Mr H Richemond

Issue No : 01
Date of Issue : 15 September 2025
Expiry Date : 14 September 2030

ITEM	MEASURED QUANTITY OR TYPE OF GAUGE OR INSTRUMENT	METHOD	RANGE OF MEASURED QUANTITY	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (±)
1	Mass pieces	Internal: *EMS/NML/01* Reference: *OIML R111-1* *SADCAS TR 15* *SADCAS TR 16*	1 g 2 g 5 g 10 g 20 g 50 g 100 g 200 g 500 g 1000 g 2000 g 5000 g 10000 g 20000 g	0,1 mg 0,1 mg 0,1 mg 0,1 mg 0,1 mg 0,1 mg 0,1 mg 12 mg 12 mg 13 mg 13 mg 30 mg 120 mg 120 mg

Original date of accreditation: 4 November 2010

The CMC, expressed as an expanded uncertainty of measurement, is stated as the standard uncertainty of measurement multiplied by a coverage factor k = 2, corresponding to a confidence level of approximately 95%.

Pinkie J Malebe
SADCAS Technical Manager